



SOLICITATION ADDENDUM

Date: 2/7/2020
Subject: 35-X162758 Vasilia Ceilometer
Solicitation Number: 6100049902
Due Date/Time: 2/26/2020 9:00 AM
Addendum Number: 1

To All Suppliers:

The Commonwealth of Pennsylvania defines a solicitation “Addendum” as an addition to or amendment of the original terms, conditions, specifications, or instructions of a procurement solicitation (e.g., Invitation for Bids or Request for Proposals).

List any and all changes:

Questions on the above solicitation:

Q: Is the intended application detection of cirrus clouds? If not, please provide application.

A: The purpose of the ceilometer is to detect the change in the mixing layer height throughout the day.

Q: Will the sensor be integrated into a larger system?

A: No, it will be stand-alone.

Type of Solicitation: Electronic Bid (SRM) - Review the Questions section of your solicitation response to ensure you have responded, as required, to any questions relevant to solicitation addenda issued subsequent to the initial advertisement of the solicitation opportunity.

Except as clarified and amended by this Addendum, the terms, conditions, specifications, and instructions of the solicitation and any previous solicitation addenda, remain as originally written.

Respectfully,

Name: Janet Noll
Title: Purchasing Specialist
Phone: 717-772-5727
Email: janoll@pa.gov

BOP-1305

Revised 05/20/2015